




**DIGITAL CONDUCTIVITY METER
(Handy Model)
DB-1045**



Conductivity Meter "Handy" Model

TECHNICAL SPECIFICATIONS

Display	:-	3½ digit LCD,
Range	:-	0-1999 us, 0-19.99 ms
Resolution	:-	1 us,
Accuracy	:-	± 1% FS,
Impedence	:-	> 10 ¹² OHMs,
Sensor	:-	Conductivity Cell (Dip Type),
Battery	:-	9V DC.



DECIBEL DYNAMICS LIMITED

Chandigarh Office :-

Plot No.-182/45, Industrial Area,
Phase - 1, Chandigarh-160002.
Phone No. :- 091-0172-2650130, 2653182,
Telefax :- 0172-2650130

Delhi Office :-

64-B-UA, Jawahar Nagar
Delhi-110007. (India).
Phone No. :- 091-011-23857091, 23857330
Telefax :- 011-23857091